

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.957 °C	9.8 °C	-0.157 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.955 °C	19.7 °C	-0.255 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.052 °C	29.6 °C	-0.452 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	19.517 %HR	20.0 %HR	0.483 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.689 %HR	47.0 %HR	-1.689 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.509 %HR	80.0 %HR	-0.509 %HR	± 5.0 %HR	± 1.9E+00 %HR