

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
10 °C	9.957 °C	9.8 °C	-0.157 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.955 °C	19.7 °C	-0.255 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.052 °C	29.6 °C	-0.452 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	19.517 %HR	20 %HR	0.483 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.689 %HR	47 %HR	-1.689 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.509 %HR	80 %HR	-0.509 %HR	± 5.0 %HR	± 1.9E+00 %HR