

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
9.9972 ml	10.0 ml	0.003 ml	± 0.06 ml	±	3.5E-01 µl
24.9590 ml	25.0 ml	0.041 ml	± 0.06 ml	±	3.5E-01 µl
50.0062 ml	50.0 ml	-0.006 ml	± 0.06 ml	±	3.5E-01 µl