

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.3445 µl	100 µl	-0.345 µl	± 0.8 µl	±	2.1E-01 µl
500.2190 µl	500 µl	-0.219 µl	± 4 µl	±	5.9E-02 µl
1000.338 µl	1000 µl	-0.338 µl	± 8 µl	±	2.3E+00 µl