

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.112 °C	30.2 °C	0.088 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	32.1 %HR	2.287 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.425 %HR	51.5 %HR	2.075 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.159 %HR	71.3 %HR	-2.859 %HR	± 3.0 %HR	± 1.9E+00 %HR