

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 1.000 mm          | 1.001 mm  | 1 µm          | ± 30 µm           | ±                           | 5.8E-01 µm |
| 1.250 mm          | 1.253 mm  | 3 µm          | ± 30 µm           | ±                           | 5.8E-01 µm |
| 1.500 mm          | 1.502 mm  | 2 µm          | ± 30 µm           | ±                           | 5.8E-01 µm |
| 2.000 mm          | 2.003 mm  | 3 µm          | ± 30 µm           | ±                           | 5.8E-01 µm |
| 3.000 mm          | 3.003 mm  | 3 µm          | ± 30 µm           | ±                           | 5.8E-01 µm |
| 4.000 mm          | 4.003 mm  | 3 µm          | ± 30 µm           | ±                           | 5.9E-01 µm |
| 5.000 mm          | 5.003 mm  | 3 µm          | ± 30 µm           | ±                           | 5.8E-01 µm |
| 8.000 mm          | 8.002 mm  | 2 µm          | ± 30 µm           | ±                           | 5.9E-01 µm |
| 10.000 mm         | 10.003 mm | 3 µm          | ± 30 µm           | ±                           | 5.8E-01 µm |