

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.000 mm	1.000 mm	0 µm	± 3 µm	±	5.8E-01 µm
1.250 mm	1.250 mm	0 µm	± 3 µm	±	5.8E-01 µm
1.500 mm	1.500 mm	0 µm	± 3 µm	±	5.8E-01 µm
2.000 mm	2.000 mm	0 µm	± 3 µm	±	5.8E-01 µm
3.000 mm	3.000 mm	0 µm	± 3 µm	±	5.8E-01 µm
4.000 mm	4.000 mm	0 µm	± 3 µm	±	5.9E-01 µm
5.000 mm	5.000 mm	0 µm	± 3 µm	±	5.8E-01 µm
8.000 mm	8.001 mm	1 µm	± 3 µm	±	5.9E-01 µm
10.000 mm	10.001 mm	1 µm	± 3 µm	±	5.8E-01 µm