

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 1.000 mm                 | 1.000 mm         | 0 µm                 | ± 30 µm                  | ±                                  | 5.8E-01 µm |
| 1.250 mm                 | 1.250 mm         | 0 µm                 | ± 30 µm                  | ±                                  | 5.8E-01 µm |
| 1.500 mm                 | 1.500 mm         | 0 µm                 | ± 30 µm                  | ±                                  | 5.8E-01 µm |
| 2.000 mm                 | 2.000 mm         | 0 µm                 | ± 30 µm                  | ±                                  | 5.8E-01 µm |
| 3.000 mm                 | 3.000 mm         | 0 µm                 | ± 30 µm                  | ±                                  | 5.8E-01 µm |
| 4.000 mm                 | 4.000 mm         | 0 µm                 | ± 30 µm                  | ±                                  | 5.9E-01 µm |
| 5.000 mm                 | 5.000 mm         | 0 µm                 | ± 30 µm                  | ±                                  | 5.8E-01 µm |
| 8.000 mm                 | 8.001 mm         | 1 µm                 | ± 30 µm                  | ±                                  | 5.9E-01 µm |
| 10.000 mm                | 10.001 mm        | 1 µm                 | ± 30 µm                  | ±                                  | 5.8E-01 µm |