

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.141 °C	0.2 °C	0.341 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.486 °C	20.5 °C	0.014 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.277 °C	34.9 °C	-0.377 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.273 %HR	33.8 %HR	3.527 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.161 %HR	52.9 %HR	-2.261 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.264 %HR	80.9 %HR	0.636 %HR	± 3.0 %HR	± 1.9E+00 %HR