

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.141 °C	0.2 °C	0.341 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.486 °C	20.6 °C	0.114 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.277 °C	34.9 °C	-0.377 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.273 %HR	31.7 %HR	1.427 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.161 %HR	54.5 %HR	-0.661 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.264 %HR	77.5 %HR	-2.764 %HR	± 3.0 %HR	± 1.9E+00 %HR