

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | -0.141 °C                  | 0.3 °C             | 0.441 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.486 °C                  | 20.2 °C            | -0.286 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 35 °C                      | 35.277 °C                  | 35.5 °C            | 0.223 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.273 %HR                 | 28.7 %HR           | -1.573 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 55 %HR                     | 55.161 %HR                 | 54.6 %HR           | -0.561 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 80 %HR                     | 80.264 %HR                 | 79.2 %HR           | -1.114 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |