

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.1440 µl	100 µl	-0.1440 µl	± 0.8 µl	±	5.3E-02 µl
499.8180 µl	500 µl	0.1820 µl	± 4 µl	±	6.1E-01 µl
1000.378 µl	1000 µl	-0.3778 µl	± 8 µl	±	7.9E-01 µl