

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
12.02 RPM	12 RPM	-0.02 RPM	± 1 RPM	±	5.8E-01 RPM
50.00 RPM	50 RPM	0.00 RPM	± 1 RPM	±	5.8E-01 RPM
100.00 RPM	100 RPM	0.00 RPM	± 1 RPM	±	5.8E-01 RPM