

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
12.04 RPM	12 RPM	-0.04 RPM	± 1 RPM	±	5.8E-01 RPM
50.04 RPM	50 RPM	-0.04 RPM	± 1 RPM	±	5.8E-01 RPM
100.02 RPM	100 RPM	-0.02 RPM	± 1 RPM	±	5.9E-01 RPM