

Certificado No. (Certificate #):58246Equipo (Instrument):Reloj Indicador

Fecha de Calibración (Calibration Date):2025-06-24Marca (Brand):Gefran

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.0000 mm	1.0300 mm	30.00 µm		±	5.8E+00 µm
1.0500 mm	1.0850 mm	35.00 µm		±	5.8E+00 µm
2.0000 mm	2.2650 mm	265.00 µm		±	5.8E+00 µm
8.0000 mm	8.2725 mm	272.50 µm		±	5.8E+00 µm
10.0000 mm	10.2600 mm	260.00 µm		±	5.8E+00 µm
15.0000 mm	15.2625 mm	262.50 µm		±	5.8E+00 µm
20.0000 mm	20.2575 mm	257.50 µm		±	5.8E+00 µm
22.0000 mm	22.2700 mm	270.00 µm		±	5.8E+00 µm
25.0000 mm	25.2500 mm	250.00 µm		±	5.8E+00 µm