

Certificado No. (Certificate #):		58084		Equipo (Instrument):		Trazador de Alturas	
Fecha de Calibración (Calibration Date):		2025-06-26		Marca (Brand):		No Indcada	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Altura (Height)		1.000 in	1.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		2.000 in	2.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		4.000 in	4.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		5.000 in	5.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		6.000 in	6.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		7.000 in	7.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		9.000 in	9.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		10.000 in	10.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		13.000 in	13.000 in	0.0000 in		±	7.2E-02 in
Altura (Height)		15.000 in	15.000 in	0.0000 in		±	7.2E-02 in