

Certificado No. (Certificate #):		58091		Equipo (Instrument):		Reloj Palpador			
Fecha de Calibración (Calibration Date):		2025-06-26		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		Exactitud (Accuracy)		Incertidumbre (Uncertainty)	
0.0500 mm		0.05 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
0.1000 mm		0.10 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
0.3000 mm		0.30 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
0.5000 mm		0.50 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
0.8000 mm		0.80 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
1.0000 mm		1.00 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
1.2000 mm		1.20 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
1.4000 mm		1.40 mm		0.00 µm		± 16 µm		±	5.8E+00 µm
1.6000 mm		1.60 mm		0.00 µm		± 16 µm		±	5.8E+00 µm