

| Certificado No. (Certificate #):         |  | 58092      |  | Equipo (Instrument): |  | Reloj Palpador       |  |                             |            |
|------------------------------------------|--|------------|--|----------------------|--|----------------------|--|-----------------------------|------------|
| Fecha de Calibración (Calibration Date): |  | 2025-06-26 |  | Marca (Brand):       |  | Mitutoyo             |  |                             |            |
| Patrón (Standard)                        |  | UBP (UUT)  |  | Error (Error)        |  | Exactitud (Accuracy) |  | Incertidumbre (Uncertainty) |            |
| 0.0500 mm                                |  | 0.05 mm    |  | 0.00 µm              |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 0.1000 mm                                |  | 0.10 mm    |  | 0.00 µm              |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 0.3000 mm                                |  | 0.30 mm    |  | 0.00 µm              |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 0.5000 mm                                |  | 0.50 mm    |  | 2.50 µm              |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 0.8000 mm                                |  | 0.81 mm    |  | 7.50 µm              |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 1.0000 mm                                |  | 1.01 mm    |  | 7.50 µm              |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 1.2000 mm                                |  | 1.21 mm    |  | 10.00 µm             |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 1.4000 mm                                |  | 1.41 mm    |  | 10.00 µm             |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |
| 1.6000 mm                                |  | 1.61 mm    |  | 10.00 µm             |  | ± 16 µm              |  | ±                           | 5.8E+00 µm |