

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.564 °C	30.5 °C	-0.064 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.273 %HR	33.0 %HR	2.727 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	51.3 %HR	1.251 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.0 %HR	-2.243 %HR	± 3.0 %HR	± 1.9E+00 %HR