

| Patrón <i>(Standard)</i> | UBP <i>(UIT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |             |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|-------------|
| 12.02 RPM                | 12 RPM           | -0.02 RPM            | ± 1 RPM                     | ±                                  | 5.8E-01 RPM |
| 50.00 RPM                | 50 RPM           | 0.00 RPM             | ± 1 RPM                     | ±                                  | 5.8E-01 RPM |
| 99.90 RPM                | 99.9 RPM         | 0.00 RPM             | ± 1 RPM                     | ±                                  | 5.8E-01 RPM |