

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
5.0122 µl	5 µl	-0.0122 µl	± 0.125 µl	±	5.1E-02 µl
25.0510 µl	25 µl	-0.0510 µl	± 0.5 µl	±	1.1E+00 µl
49.9217 µl	50 µl	0.0783 µl	± 0.5 µl	±	4.0E-01 µl