

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
5.01221 µl	5 µl	-0.01221 µl	± 0.125 µl	±	5.1E-02 µl
24.97085 µl	25 µl	0.02915 µl	± 0.5 µl	±	5.1E-02 µl
50.12214 µl	50 µl	-0.12214 µl	± 0.5 µl	±	5.2E-02 µl