

| Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|----------------------|-----------------------------|------------|
| 5.01221 µl        | # 5 µl    | -0.01221 µl   | ± 0.125 µl           | ±                           | 5.1E-02 µl |
| 24.97085 µl       | 25 µl     | 0.02915 µl    | ± 0.5 µl             | ±                           | 5.1E-02 µl |
| 50.12214 µl       | 50 µl     | -0.12214 µl   | ± 0.5 µl             | ±                           | 5.2E-02 µl |