

Certificado No. (Certificate #):58136.1Equipo (Instrument):Medidor de Espesores

Fecha de Calibración (Calibration Date):2025-07-08Marca (Brand):Defelsko

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
25.0000 μm	27.2500 μm	2.2500 μm		±	9.4E-01 μm
52.0000 μm	50.5000 μm	-1.5000 μm		±	7.4E-01 μm
54.0000 μm	48.0000 μm	-6.0000 μm		±	1.4E+00 μm
129.0000 μm	132.5000 μm	3.5000 μm		±	1.4E+00 μm
258.0000 μm	248.0000 μm	-10.0000 μm		±	7.0E+00 μm
503.0000 μm	501.0000 μm	-2.0000 μm		±	2.5E+01 μm
511.0000 μm	496.5000 μm	-14.5000 μm		±	1.3E+01 μm