

Certificado No. (Certificate #):58136.2Equipo (Instrument):Medidor de Espesores

Fecha de Calibración (Calibration Date):2025-07-08Marca (Brand):Defelsko

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
25.0000 μm	24.7500 μm	-0.2500 μm		±	2.9E+00 μm
52.0000 μm	52.0000 μm	0.0000 μm		±	5.8E-01 μm
54.0000 μm	53.5000 μm	-0.5000 μm		±	3.5E+00 μm
129.0000 μm	127.0000 μm	-2.0000 μm		±	5.8E-01 μm
258.0000 μm	245.0000 μm	-13.0000 μm		±	2.4E+00 μm
503.0000 μm	490.0000 μm	-13.0000 μm		±	5.8E-01 μm
511.0000 μm	500.0000 μm	-11.0000 μm		±	5.8E-01 μm