

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
50.2 mmHg	50 mmHg	-0.2 mmHg	± 3 mmHg	±	1.2E+00 mmHg
100.4 mmHg	100 mmHg	-0.4 mmHg	± 3 mmHg	±	1.2E+00 mmHg
150.4 mmHg	150 mmHg	-0.4 mmHg	± 3 mmHg	±	1.2E+00 mmHg
200.4 mmHg	200 mmHg	-0.4 mmHg	± 3 mmHg	±	1.2E+00 mmHg
250.6 mmHg	250 mmHg	-0.6 mmHg	± 3 mmHg	±	1.2E+00 mmHg
300.8 mmHg	300 mmHg	-0.8 mmHg	± 3 mmHg	±	1.2E+00 mmHg