

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.402 °C	16.8 °C	1.398 °C	± 1.0 °C	± 4.0E-01 °C
23 °C	23.173 °C	23.5 °C	0.327 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.082 °C	29.4 °C	-0.682 °C	± 1.0 °C	± 4.0E-01 °C
70 %HR	70.120 %HR	67 %HR	-3.120 %HR	± 5.0 %HR	± 1.9E+00 %HR