

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60 RPM	60 RPM	0 RPM	± 0.012 RPM	±	5.8E-02 RPM
600 RPM	600 RPM	0 RPM	± 0.12 RPM	±	5.8E-02 RPM
6000 RPM	6000 RPM	0 RPM	± 1.2 RPM	±	5.8E-02 RPM
60000 RPM	60001 RPM	1 RPM	± 12 RPM	±	8.0E-01 RPM
99960 RPM	99962 RPM	2 RPM	± 20 RPM	±	8.4E-01 RPM