

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
50.4 mmHg	50 mmHg	-0.4 mmHg	± 3 mmHg	±	1.2E+00 mmHg
100.2 mmHg	100 mmHg	-0.2 mmHg	± 3 mmHg	±	1.2E+00 mmHg
150.6 mmHg	150 mmHg	-0.6 mmHg	± 3 mmHg	±	1.2E+00 mmHg
200.6 mmHg	200 mmHg	-0.6 mmHg	± 3 mmHg	±	1.2E+00 mmHg
250.7 mmHg	250 mmHg	-0.7 mmHg	± 3 mmHg	±	1.2E+00 mmHg
301.2 mmHg	300 mmHg	-1.2 mmHg	± 3 mmHg	±	1.2E+00 mmHg