

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1 mm	0.98 mm	-20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	1.50 mm	1.48 mm	-20 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	3.0 mm	2.98 mm	-20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	5 mm	4.98 mm	-20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	8 mm	7.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	10 mm	9.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	15 mm	14.98 mm	-20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	20 mm	19.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	25 mm	24.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm