

Certificado No. (Certificate #): 58169

Equipo (Instrument): Micrómetro de Exterior

Fecha de Calibración (Calibration Date): 2025-06-20

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
0.0500 in	0.0500 in	0.0000 in	± 0.0001 in	± 3.2E-05 in
0.1000 in	0.1000 in	0.0000 in	± 0.0001 in	± 3.3E-05 in
0.1010 in	0.1010 in	0.0000 in	± 0.0001 in	± 3.3E-05 in
0.1050 in	0.1050 in	0.0000 in	± 0.0001 in	± 3.2E-05 in
0.1100 in	0.1100 in	0.0000 in	± 0.0001 in	± 3.2E-05 in
0.1500 in	0.1500 in	0.0000 in	± 0.0001 in	± 3.4E-05 in
0.2000 in	0.2000 in	0.0000 in	± 0.0001 in	± 3.3E-05 in
0.3000 in	0.3000 in	0.0000 in	± 0.0001 in	± 3.3E-05 in
0.4000 in	0.4000 in	0.0000 in	± 0.0001 in	± 2.9E-05 in
0.5000 in	0.5000 in	0.0000 in	± 0.0001 in	± 3.1E-05 in
1.0000 in	1.0000 in	0.0000 in	± 0.0001 in	± 3.3E-05 in