

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.263 °C	30.1 °C	-0.163 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.171 %HR	35.1 %HR	-0.121 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	50.8 %HR	0.782 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.7 %HR	1.252 %HR	± 3.0 %HR	± 1.9E+00 %HR