

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.402 °C	15.7 °C	0.298 °C	± 1.0 °C	± 4.0E-01 °C
23 °C	23.173 °C	23.1 °C	-0.073 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.082 °C	29.8 °C	-0.282 °C	± 1.0 °C	± 4.0E-01 °C
70 %HR	70.120 %HR	65 %HR	-5.120 %HR	± 5.0 %HR	± 1.9E+00 %HR