

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3600.132 s	3600 s	-0.132 s	-0.0037%	± 0.0556%	±	5.6E-02 s
3600.146 s	3600 s	-0.146 s	-0.0041%	± 0.0556%	±	5.6E-02 s
3600.289 s	3600 s	-0.289 s	-0.0080%	± 0.0556%	±	5.6E-02 s