

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	15.5 °C	0.479 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.018 °C	24.8 °C	-0.218 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.017 °C	50.8 °C	0.783 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.028 %HR	29.0 %HR	3.972 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	46.0 %HR	-4.263 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	71.0 %HR	-4.560 %HR	± 5.0 %HR	± 1.9E+00 %HR