

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.021 °C	15.9 °C	0.879 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.018 °C	25.5 °C	0.482 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.017 °C	49.2 °C	-0.817 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.028 %HR	28.0 %HR	2.972 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	47.0 %HR	-3.263 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	71.0 %HR	-4.560 %HR	± 5.0 %HR	± 1.9E+00 %HR