

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.965 °C	2.7 °C	0.735 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.082 °C	8.3 °C	0.218 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.500 °C	25.3 °C	-0.200 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.223 °C	29.7 °C	-0.523 °C	± 1.0 °C	± 4.0E-01 °C
70 %HR	70.918 %HR	68.0 %HR	-2.918 %HR	± 4.0 %HR	± 1.9E+00 %HR