

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	31.055 °C	31 °C	-0.055 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.668 %HR	28 %HR	-0.668 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.721 %HR	48 %HR	-1.721 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.593 %HR	70 %HR	-1.593 %HR	± 3.0 %HR	± 1.9E+00 %HR