

Certificado No. (Certificate #):58246Equipo (Instrument):Reloj Indicador

Fecha de Calibración (Calibration Date):2025-06-24Marca (Brand):Gefran

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.0000 mm	1.0400 mm	40.0000 µm		±	5.8E+00 µm
1.0500 mm	1.0950 mm	45.0000 µm		±	5.8E+00 µm
2.0000 mm	2.2750 mm	275.0000 µm		±	5.8E+00 µm
8.0000 mm	8.2825 mm	282.5000 µm		±	5.8E+00 µm
10.0000 mm	10.2700 mm	270.0000 µm		±	5.8E+00 µm
15.0000 mm	15.2725 mm	272.5000 µm		±	5.8E+00 µm
20.0000 mm	20.2675 mm	267.5000 µm		±	5.8E+00 µm
22.0000 mm	22.2850 mm	285.0000 µm		±	5.8E+00 µm
25.0000 mm	25.2767 mm	276.6667 µm		±	5.8E+00 µm