

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
130.3 μm	131.8 μm	1.5 μm	± 5 μm	±	5.8E-01 μm
523.0 μm	522.6 μm	-0.4 μm	± 5 μm	±	5.8E-01 μm