

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 52.65 μm          | 52.74 μm  | 0.09 μm       | ± 2.5 μm          | ±                           | 5.8E-01 μm |
| 155.23 μm         | 155.4 μm  | 0.17 μm       | ± 2.5 μm          | ±                           | 5.8E-01 μm |
| 257.12 μm         | 527.4 μm  | 270.28 μm     | ± 2.5 μm          | ±                           | 5.8E-01 μm |
| 511.72 μm         | 512.2 μm  | 0.48 μm       | ± 2.5 μm          | ±                           | 5.8E-01 μm |