

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.764 °C	1.8 °C	0.036 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.760 °C	4.0 °C	0.240 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.691 °C	7.8 °C	0.109 °C	± 0.5 °C	± 4.0E-01 °C