

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.764 °C	2.0 °C	0.236 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.760 °C	4.1 °C	0.340 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.691 °C	7.9 °C	0.209 °C	± 0.5 °C	± 4.0E-01 °C