

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.764 °C	1.8 °C	0.036 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.760 °C	3.9 °C	0.140 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.691 °C	7.7 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C