

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3599.567 s	3599.50 s	-0.067 s	-0.002%	± 0.0012 %	±	5.6E-02 s
3599.247 s	3599.31 s	0.063 s	0.002%	± 0.0012 %	±	5.6E-02 s
3599.459 s	3599.50 s	0.041 s	0.001%	± 0.0012 %	±	5.6E-02 s