

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	10.4 °C	0.393 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	24.9 °C	-0.449 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	32.4 °C	0.332 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	32.7 %HR	2.539 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.501 %HR	69.9 %HR	-0.601 %HR	± 3.0 %HR	± 1.9E+00 %HR