

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | -0.893 °C                  | 0.7 °C             | 1.593 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 15 °C                      | 14.931 °C                  | 15.6 °C            | 0.669 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 29.641 °C                  | 29.9 °C            | 0.259 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 28.586 %HR                 | 30 %HR             | 1.414 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 71.593 %HR                 | 69 %HR             | -2.593 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |