

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.893 °C	0.9 °C	1.793 °C	± 1.0 °C	± 4.0E-01 °C
15 °C	14.931 °C	16.6 °C	1.669 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	29.641 °C	30.4 °C	0.759 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.586 %HR	24.6 %HR	-3.986 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.593 %HR	63.9 %HR	-7.693 %HR	± 5.0 %HR	± 1.9E+00 %HR