

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.00920 in	0.0093 in	0.0001 in	± 0.0002 in	±	5.9E-05 in
0.00203 in	0.0020 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
0.00491 in	0.0049 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
0.00985 in	0.0099 in	0.0001 in	± 0.0002 in	±	5.8E-05 in
0.01913 in	0.0191 in	0.0000 in	± 0.0002 in	±	6.0E-05 in
0.03791 in	0.0379 in	0.0000 in	± 0.0002 in	±	6.4E-05 in
0.0625 in	0.0631 in	0.0006 in	± 0.0002 in	±	7.3E-05 in
0.100 in	0.1007 in	0.0007 in	± 0.0002 in	±	9.1E-05 in
0.200 in	0.2008 in	0.0008 in	± 0.0002 in	±	1.5E-04 in