

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.007 °C	10.7 °C	0.693 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.2 °C	0.179 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.2 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	32.3 °C	0.232 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	32.2 %HR	2.039 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.501 %HR	68.5 %HR	-2.001 %HR	± 3.0 %HR	± 1.9E+00 %HR